

Calibration Certificate

Bartington Instruments Ltd, 10 Thorney Leys Business Park, Witney, Oxon. OX8 7GE, UK

The following measurements were carried out using working standards which are periodically verified, and are traceable to National Physical Laboratory Standards.

Calibration References

Bartington Mag-01 Magnetometer	TE112	Serial No 0288
Bartington Mag Probe B	TE113	Serial No 501
Bartington Solenoid with Current Source	TE18	Serial No 001

Three axis magnetic field sensor under test

Mag-03 *MSES* Range *100 µT* Serial No *651*

Test Results

PARAMETER	SPECIFIED		RESULTS		
			X	Y	Z
Orthogonality Error (°)	MS and MSES versions ±0.1° All other versions ±0.5°		✓	✓	✓
Offset Error in Zero Field (nT)	Range (µT)	Offset (nT)			
	70	±5			
	100	±5	✓	✓	✓
	250	±12			
	500	±25			
	1000	±50			
Scaling Error* (@35Hz) (%)	± 0.5 %		0%	0%	-0.2%
Frequency Response** (%)	± 5 %		-2.8%	-4.0%	-3.8%
Noise at 1 Hz (pTrms/√Hz)	Standard - 12 pTrms/√Hz Low noise - 6 pTrms/√Hz		4.8	4.6	4.3

* Measured at 35 Hz using a field of 100 µT pk-pk.

** The response at 1 kHz referred to that at 35 Hz.

The measurement uncertainty for the above values is ± 1 LSD

The calibration was measured at 20 (± 2) °C

Certified that the above instrument has been fully tested and calibrated to the manufacturers published specifications, due allowance having been made for the uncertainty of the measurements.

NAME..... *E. McLACHLAN* Date..... *8/5/02*

Signature..... *E. McLachlan*